Technical Service Consultants Ltd

Microbiology House, Fir Street, Heywood, Lancashire OL10 1NW, UK T : +44 (0)1706 620600 $\,$ F : +44 (0)1706 620445

E: sales@tscswabs.co.uk W: www.tscswabs.co.uk

Product Specification Sheet

TS/15-F Sterile 500 ml Empty Sampling Container

| Product Code | TS/15-F |
|-----------------------------|-----------------------------------------------------------------------------------------------------------------------|
| Description | Sterile 500 ml Empty Sampling Container |
| Intended Use | Sampling of potable and swimming pool water samples for microbiological analysis or for general purpose applications. |
| Components | HDPE Bottle HDPE Cap Label |
| Contents | Not applicable as supplied empty. |
| Method of Sterilisation | Sterilised by gamma irradiation at dosage of 25-46kGy |
| Shelf Life | 24 months from the date of manufacture |
| Storage Conditions | 4-25 °C. Store in dry and ventilated area; protect from light, heat and moisture in tightly closed original container |
| Precautions and Limitations | |
| Pack Size | 100 in a box |
| Other | Quality control testing details on final product:1) Validation of irradiation sterility |

This specification is intended for guidance only. You should always refer to a professional with experience in your intended application if in any doubt.

Issue #: 3 Date: 06.06.2016 ProductSpec_TS15-F